

DBH
TLW

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/15/2015

Instrument Serial Number:	010596	As of 15-Apr-15
---------------------------	--------	-----------------

Start Date	End Date	Agency Name
14-Apr-15		DFS Central Lab
28-Oct-14	14-Apr-15	Appomattox County SO
21-Oct-14	28-Oct-14	DFS Central Lab
23-May-14	21-Oct-14	West Point PD
21-May-14	23-May-14	DFS Central Lab
13-Nov-08	21-May-14	Bath County SO
15-Aug-08	13-Nov-08	DFS Central Lab

DBH
THW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010596 Worksheet Start Date 4/14/2015
Location Appomattox CO SO
Address 179 Morton Lane Appomattox VA 24522
DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 739 Reference Barometer (mm HG) 742
Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.299
0.300	0.292	0.283	0.300	Sample 2	0.299
Precision		sample min	sample max	Sample 3	0.298
0.001		0.298	0.299		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.097	0.094	0.100	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		


Dry gas standard Lot No. (with tank no.) AG425201-17

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010596

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

4/15/15

Issuing Analyst

T. [Signature]

Date

*4/16/15**JPD*

DBH
TLN




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/14/2015
TEST EQUIPMENT NUMBER 010596		

RESULTS: TIME SAMPLE TAKEN 10:23 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 15 Apr 2015 at 13:45

DBH
TLN

Test Results

Instrument Serial Number 010596

Test # 000348 Subject Test

Test Location 1 Department of
Test Date 14 Apr 2015

Test Location 2 Forensic Science
Test Time 10:16
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B
License Number 21250

Card Serial Number 121250
Subject's Last Name INSTRUMENT

Effective Date 10/01/2014 Expiration Date 10/01/2016

Subject's Middle Initial

Subject's First Name TEST

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 14 Apr 2015 End Time 10:24

Result Time 10:23

Result Date 14 Apr 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:17

Data Type BLK

Sample Value 0.000

Sample Time 10:17

Data Type CHK

Sample Value 0.099

Sample Time 10:18

Data Type BLK

Sample Value 0.000

Sample Time 10:19

Data Type SUBJ

Sample Value 0.000

Sample Time 10:20

Data Type BLK

Sample Value 0.000

Sample Time 10:21

Data Type SUBJ

Sample Value 0.000

Sample Time 10:23

Data Type BLK

Sample Value 0.000

Sample Time 10:24

Standard Type Dry Gas Std

Standard Value 0.097

Standard Lot Number AG425201-17

Standard Expiration Date 09/09/2016

Tank Pressure 1071

Barometric Pressure 739 mmHg

Blow Sample Number 1 Blow Duration 3.79 sec

Blow Volume 1914 cc End-of-Blow Time 10:20

Blow Sample Number 2 Blow Duration 3.90 sec

Blow Volume 1903 cc End-of-Blow Time 10:23

Tamper Evident Stamp 526c5593

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

DBH
TL

Department of Forensic Science

Serial Number: 010596 Test Number: 349

Test Date: 04/14/2015 Test Time: 10:25 EDT

Dry Gas Target: 0.097

Lot Number: AG425201-17 Exp Date: 09/09/2016

Tank Pressure: 1066 psi Barometric Pressure: 739 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:26
CHK	0.099	10:27
BLK	0.000	10:28
CHK	0.099	10:29
BLK	0.000	10:30
CHK	0.099	10:31

Test Status: *Success*

Calibration CRC: 2935AE90

Intox EC/IR-II: Accuracy Check

DBH
TLV

Department of Forensic Science

Serial Number: 010596 Test Number: 350

Test Date: 04/14/2015 Test Time: 10:37 EDT

Dry Gas Target: 0.292

Lot Number: AG428003-05 Exp Date: 10/07/2016

Tank Pressure: 232 psi Barometric Pressure: 739 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	10:38
CHK	0.299	10:39
BLK	0.000	10:40
CHK	0.299	10:41
BLK	0.000	10:43
CHK	0.298	10:44

Test Status: Success

Calibration CRC: 2935AE90